



OFFICE OF THE DIRECTOR ROURKELA GOVT. HOSPITAL



E-Mail: directorrrghrourkela@gmail.com, Phone/Fax: (0661) 2400161/240003

Quotation No: 6086 /RGH

Date: 03.12.2025


QUOTATION CALL NOTICE

Sealed quotations are invited from authorized interested bidders/suppliers to quote their lowest rate for "Supply of RF-Blood Bag Tube Sealer for Blood Bank, RGH Rourkela. The quotation is to be sealed and super scribed "Quotation for Supply of RF-Blood Bag Tube Sealer to Blood Bank, RGH Rourkela" the details specification and description is as below.

<u>Sl. No</u>	<u>Specification</u>	<u>Unit</u>	<u>Amount (Rs)</u>
1.	<p><u>Table Top Tube Sealer (RF-Blood Bag Tube Sealer) :</u></p> <ol style="list-style-type: none"> It should be user friendly with power backup system. It should have multi functional Indicator to guide user. It should have easily accessible electrode, for easy cleaning and disinfection Its electrode should have a hermetic sealing through high radio frequency (RF) waves. (Up to 40.68 mhz) Importance is it heat only the tube not the liquid inside and thereby reduces hemolysis of blood and blood components). Its electrode should have splash guard to protect user. Its electrode should automatically detect blood bag tube. (PVC). It sealing time should be 1- 3 second with provision for adjustable mode. It should be IEC 61010-1:2010 safety complied with protection against electric shock. It should work on power supply 100 V AC to 240 V AC 50 Hz/60 Hz with Power consumption range Operating: 185 W Standby: 20 W. It should be internationally complied to CE mark under Low Voltage Directive (2014/35/EU), RoHS 3 (EU Directive 2015/863) It should be table top model with weigh within 5-6 kg 	Per Item	

Note Below:

- Bidder/Agency should have submitted valid PAN and GST
- Quotation will be accepted through Indian Post/Couriers.
- No Quotation will be accepted after stipulated date and time i.e., 10.12.2025 at 3 pm.
- Quotation will be opened in the stipulated date and time i.e., 10.12.2025 at 5 pm.
- The authorities reserve the right to cancel all quotations without assigning any reason thereof.


Director I/c 03/12/25

Rourkela Govt. Hospital





OFFICE OF THE DIRECTOR

ROURKELA GOVT. HOSPITAL

E-Mail: directorrrghrourkela@gmail.com, Phone/Fax: (0661) 2400161/240003




Memo No:

6007 (3)

Date:

03-12-2025

1. Copy forwarded to the **General Manager, NIC, Sundargarh** district for kind information and a request to publish at district website for wide publication.
2. Copy submitted to the **Collector & DM, Sundargarh** for favour of kind information and necessary action.



Director I/c
Rourkela Govt. Hospital